

**Product Specification Sheet** 


# POZ-TX Percent Oxygen Sensor



### **Specifications:**

Micro Fuel Cell
245 - 430 uA
0 - 100 Percent
7 Seconds
± 1% of Signal
± 0.5%
2.54 % / ° C
0 to 50 ° C
0 to 45 ° C
0.5 - 2 SCFH
0 - 100% RH (Non-Condensing)
12 - 15 Months
6 Months
9 Months
CenterFoilNegative Outer Foil Positive


- 1. Signal output is measured in air at 25  $^{\circ}$  C, sea level.
- 2. Accuracy full scale is calculated with constant pressure and temperature & proper calibration (90% O2 value on full scale range). Drastic temperature change can result in a maximum error of  $\pm$  10%.
- 3. Expected life is calculated when O2 is 20.9% @ 25 ° C, sea level.
- 4. Southland Sensing warrants these sensors for the period noted above to be free from defects in materials and workmanship. Southland Sensing will not be held liable for sensors damaged due to customer neglect.



#### **Sensing Surface**



#### **PCB** Assembly



## Direct Replacement Guide

Analytical Industries: PSR-11-21

Teledyne: B-1

#### 深圳市新世联科技有限公司

地址:深圳市深南中路2066号华能大厦712室

电话: 0755-83680810 83680820 83680830 83680860

网址: www.apollounion.com

邮编: 518031 传育: 0755-83680

传真: 0755-83680866 邮箱: sales@apollounion.com